




AE176

Blood Pressure Monitor


GETTING STARTED

GET STARTED

Install 4 AAA batteries or connect Micro USB cable to a 5V powered USB port.



Plug arm cuff into the main device.



SYSTEM SETTING

Press the **SET button** and hold it for more than **3 SECONDS**. The User symbol will then flash. Set User, Date and Time, and Talking Volume follow the steps below.

A. Setting the User

Press the **MEM button** to select User A or User B. Then press **SET button** to confirm.


B. Setting Date and Time

When the year display is flashing, press the **MEM button** to advance to the current year, then press **SET button** to confirm. Change month, date and time following the same steps.

C. Setting Talking Volume

When the SP display is flashing, press **MEM button** to select talking volume 1 (Low), volume 2 (Medium), volume 3 (High) or OFF. Press **SET button** to confirm.

After finish setting for one user, set up the other user following the same steps.



Place cuff directly on bare skin of upper left arm, with the bottom of the cuff edge **2-3 CM** above elbow.


Align cuff so the air tube is centered on the inside of your arm. Wrap cuff snugly.



Press **START/STOP button** to start measurement.

The cuff will inflate automatically. Hold still until the cuff deflates and results are displayed.





SWITCH USERS

Press SET button and hold it for more than 3 seconds until the User Symbol flash. Press MEM button to select user A or B. Then press START/STOP button to turn off the device.



VIEW THE MEASUREMENT VALUES IN MEMORY

Select User ID (A or B)

Follow instructions on SWITCH USERS section.

Press MEM Button

The first reading that appears is the Average value for last 3 readings. Press **MEM button** to view the values stored in memory.

MUTE/UNMUTE

Step 1

When the monitor is **OFF**, press the **SET** button. The time and user symbol will display.

Step 2

Press **SET** button to turn on the mute.

Step 3

Press the **SET** button again to turn off the mute.





FREQUENTLY ASKED QUESTIONS

Q. Why are my blood pressure readings sometimes different?

A. People's blood pressure could change constantly throughout the day and night, or in different conditions when you measure your blood pressure. Many factors, such as activity level, food/drink intake, stress can affect a person's blood pressure. Because of this, the readings may fluctuate. That's why it is important to use averages over time to compare readings and review the overall results with your doctor.

Q. Why could my doctor's readings be different than my readings taken at home?

A. Blood pressure readings measured by a healthcare professional in a doctor's office are usually somewhat higher than that you measure it yourself in a relaxed atmosphere at home. We call this "doctor's surgery" or "white coat" hypertension. Because these readings can vary considerably, you should take your blood pressure on a regular basis at home and share your results with your healthcare provider.

OTHER IMPORTANT TIPS

To obtain more accurate readings, sit quietly with feet flat on the floor and rest for 15 minutes before taking a measurement. Do not measure within 30 minutes after eating, exercising, bathing, smoking or drinking alcohol.

The Blood Pressure Indicator show how your reading compares to the internationally recognized guidelines for normal home blood pressure levels.

This monitor allows 2 users to monitor and track their readings separately in the stored memory. To review readings for multiple users, select your USER ID (A or B).

EXTENDED WARRANTY

Go to www.AlcedoHealth.com to register your product and receive an extended warranty.

Please read the instruction manual for complete product information.

Have questions?

DO NOT RETURN THIS PRODUCT.

Call our Toll-Free Consumer Help Line at 833-682-8655 or email info@alcedohealth.com for assistance with your Alcedo product.

Please go to our YouTube channel [Alcedo Health] for instructional video.